2018

Introduction to Laser Induced Breakdown Spectroscopy Imaging in Food: Salt Diffusion in Meat

Yash Dixit

Technological University Dublin, yash.dixit@mydit.ie

Patrick Cullen

Technological University Dublin, pj.cullen@tudublin.ie

Carl Sullivan

Technological University Dublin, carl.sullivan@tudublin.ie

Follow this and additional works at: <https://arrow.tudublin.ie/schfsehart>

Part of the [Medicine and Health Sciences Commons](#)

Recommended Citation

Dixit, Y., Cullen, P.J. & Sullivan, C. (2018). Introduction to laser induced breakdown spectroscopy imaging in food: salt diffusion in meat. *Journal of Food Engineering*, vol. 216, pp. 120-124. doi: 10.1016/j.jfoodeng.2017.08.010.

This Article is brought to you for free and open access by the School of Food Science and Environmental Health at ARROW@TU Dublin. It has been accepted for inclusion in Articles by an authorized administrator of ARROW@TU Dublin. For more information, please contact arrow.admin@tudublin.ie, aisling.coyne@tudublin.ie, gerard.connolly@tudublin.ie.

Introduction to laser induced breakdown spectroscopy imaging in food: salt diffusion in meat

Y. Dixit^{a, *}, Maria P. Casado-Gavalda ^a, R. Cama-Moncunill^a, X. Cama-Moncunill^a, Maria Markiewicz-Keszycka ^a, Franklyn Jacoby ^b, P. J. Cullen ^{a,c}, Carl Sullivan^a

^a School of Food Science and Environmental Health, Dublin Institute of Technology, Dublin 1, Ireland.

^b School of Culinary Arts and Food Technology, Dublin Institute of Technology, Dublin 1, Ireland.

^c Department of Chemical and Environmental Engineering, University of Nottingham, UK

* Corresponding author:

Y. Dixit

School of Food Science and Environmental Health,

Dublin Institute of Technology, Cathal Brugha St, Dublin 1, Ireland

Tel: +353 1 402 4543

E-mail: yash.dixit@mydit.ie

Keywords: LIBS, beef, brine, imaging, spatial.

Abstract.

This study illustrates the ability of laser induced breakdown spectroscopy (LIBS) to detect and map minerals in food. A LIBS system was used to spatially collect spectra of beef samples. Samples were brined in a 6% salt solution for 2h and 24h along with a control sample. Samples were measured by scanning the cross-section of each sample in a 90 X 90 square grid. Sodium (Na) distribution images with respect to emission peak at 589.05 nm were generated after pre-processing the spectral data which directly corresponds to salt levels. As expected, the control sample showed the lowest Na distribution whereas 2h brined sample showed distribution along the sample's edges decreasing towards the centre. The 24 h brined sample showed increased diffusion. Overall, results show the ability of LIBS to map salt diffusion in meat via Na LIBS imaging, which could be used to optimize brining conditions.

1. Introduction

Salt curing is an ancient technique which has been used over centuries for preserving perishable food products such as meat. Salt curing can be of the following types: dry, wet and a combination of both. Sodium chloride (NaCl) is the main ingredient used in meat curing. It offers various functionalities such as preservation, improved technological yields, and influences meat tissue properties such as water-holding capacity and protein solubilisation, which in turn defines meat texture (Sharedeh et al., 2015). NaCl diffusion in meat during wet (immersion) curing, also known as brining, is defined by Fick's diffusion theory. The theory of diffusion in isotropic substances is based on the hypothesis that the rate of transfer of diffusing substance through unit area of a section is proportional to the concentration gradient measured normal to the section (Crank, 1979). NaCl diffusion into the muscles is usually rapid and an equilibrium is reached in about 48 h. However, salt diffusion is slower in meats with a close tissue micro-structure when immersed in a weak brine solution (Lawrie, 2006). It is important to study salt diffusion in meat in order to optimize brining time, brine concentration as well brine temperature, which can have a direct effect on the microbiological, physico-chemical and sensorial characteristics of meat.

Laser-induced breakdown spectroscopy (LIBS) is an emerging elemental technique for mineral analysis of food (Andersen et al., 2016; Bilge et al., 2016b; Cama-Moncunill et al., 2017; Casado-Gavalda et al., 2017; Dixit et al., 2017b). LIBS provides numerous advantages such as minimal or no sample preparation, chemical free, rapid detection, portability and spatial information (Abdel-Salam et al., 2017; Bilge et al., 2016a; Er et al., 2016; Moncayo et al., 2016; Singh et al., 2017; Wang et al., 2016).

Various studies have been conducted regarding NaCl diffusion in meat during immersion-brining with or without other additives (Graiver et al., 2006, 2009; Hansen et al., 2008). However, the literature reveals no study has been reported using LIBS for analysing salt diffusion in meat. In the current study, an experiment was conducted for imaging salt diffusion in beef using a LIBS system combined with an automatic sample chamber. Chlorine emission peaks are not easily resolvable under normal atmospheric conditions due to interference with nitrogen (N) and oxygen (O) peaks, thus emission peaks related to sodium (Na) were utilized to create salt diffusion images (Weritz et al., 2007). The aim of this study was to illustrate the ability of LIBS as a novel chemical free technique for imaging NaCl diffusion in beef.

63 2. Materials and methods

64 2.1 Sample Preparation

65 Fresh round beef steak weighing approximately 1.5 Kg was purchased from a local butchers shop
66 in Dublin city and transferred to the School of Food Science and Environmental Health in the Dublin
67 Institute of Technology, Dublin, Ireland. On the same day, the steak was carefully cut in order to
68 separate the lean from the fat beef trimmings. The lean beef was cut into three cubes of approximately 4
69 cm diameter. A 6 % salt brine solution was prepared using laboratory grade NaCl (cas no: 7647-14-5,
70 >99.5 %, Sigma Aldrich, Inc.). Three beakers of volume 250 ml were used. One beaker was filled with
71 distilled water while the remaining were filled with a 6 % brine solution. The beef cubes were separately
72 immersed in beakers and were placed on a clean plastic stand to ensure a uniform diffusion from all
73 sides. Beakers were covered with aluminium foil to minimize evaporation. The sample immersed in
74 distilled water was used as a control (0 hrs). The samples in the brine solution were allowed to stand for
75 2 hrs and 24 hrs respectively. Firstly, the control sample was removed from the distilled water solution,
76 dried from all sides, using ashless paper, in order to remove any excess of liquid solution and then
77 horizontally cut to reveal the cross-section. The obtained cross-section was then subjected to LIBS
78 analysis. A similar procedure was conducted for cross-sections of the 2 hr and 24 hr brined samples.

79 2.2 LIBS spectra acquisition

80 LIBS spectra were recorded with a LIBSCAN 150 system (Applied Photonics Limited, Skipton
81 North Yorkshire, United Kingdom) used in the study by Dixit et al. (2017a). The System consists of a
82 Q-switched Nd:YAG laser (ultra, Quantel laser, 601 Haggerty Lane Bozeman, MT, USA) and a series
83 of six spectrophotometers covering the wavelength range of 185-904 nm. The head incorporates a
84 miniature CCD camera and 6 lens holders which collect plasma light of different wavelength regions.
85 The laser used for sample ablation had a pulse energy of 150 mJ and a pulse duration of 5 ns operating
86 at 1064 nm. A repetition rate of 1 Hz was employed along with a 1.27 μ s gate delay and 1.1 ms
87 integration time in Q-switched mode. The spectrograph was externally triggered from the laser at every
88 pulse with a delay generator. The sample was placed at a LTSD (lens to sample distance) of
89 approximately 80 mm to ensure that the laser was focussed onto the sample. The control sample (0 hr)
90 and 2 hr brined sample were analysed on the same day of sample preparation while the 24 hr brined
91 sample was analysed on the subsequent day. Samples were measured by applying shots in a whole
92 cross-section in a 90 X 90 square grid pattern while moving the sample after each shot by an automated
93 sample chamber (XYZ-750, Applied Photonics Limited, Skipton North Yorkshire, United Kingdom)

94 with a step size of 0.50 mm. The number of shots were selected to ensure complete coverage of the
95 sample cross-section.

96 2.3 Data analysis

97 Data analysis was performed using R (R Core Team, 2014). The packages *baseline* and *EBImage* were
98 utilized to perform spectral pre-processing and image processing respectively.

99 Initially, the spectra recorded for each sample at 90 X 90 different locations were pre-processed
100 in order to remove non-linearities introduced by light scattering, which can have a considerable effect
101 on the spectra (Rinnan et al., 2009). Pre-processing was performed using the *baseline* function in order
102 to remove background effects from the acquired spectra. The processed data obtained for the three
103 samples were transformed into individual hypercubes (Dixit et al., 2014; Gowen, 2014), where x-axis
104 and y-axis represent the coordinates (location) of the laser shots and z-axis represents the wavelength
105 range used. Hypercubes were further processed using the package *EBImage* for performing
106 morphological operations by utilizing functions “*makeBrush*” and “*filter2*” in order to obtain
107 smoothened images. Images were generated with respect to main elemental peaks using a false colour
108 scheme indicating emission intensity. Images of the emission intensity distribution for potassium (K) at
109 766.458 nm showed the maximum contrast between the background and the meat samples (Fig. 1) and
110 hence, the K emission peak at 766.458 nm was used to obtain a threshold for distinguishing the meat
111 samples from the background. It can be seen from Figure 2, this threshold performed well. It is evident
112 from Figure 1 that a correlation exists between Na (589.05 nm) and K (766.458 nm) distribution which
113 could be related to Na-K pump (Skou, 1988) and hence was utilized to normalize the spectral data. In a
114 next step, masking was performed followed by normalization using the potassium (K) emission peak at
115 766.458 nm, where the baseline corrected data of the Na emission peak (589.05 nm) intensities at each
116 shot/location were divided by the corresponding emission intensities of the K (766.458 nm) peak at the
117 same shot/location. Normalization was used in order to compensate for signal variations and sample
118 matrix differences (Castro and Pereira-Filho, 2016). Finally, salt (with regard to Na) distribution images
119 were generated. In order to compare the samples, the same intensity scale was implemented for the three
120 samples. Maximum and minimum intensities of Na emission peak (589.05 nm) for 24h brined sample
121 were used for rescaling.

3. Results and Discussion

3.1 Image analysis

Fig. 3 shows the Na distribution for the cross-section of the control, 2h brined and 24h brined samples with respect to the Na peak at 589.05 nm which directly represents the salt distribution. The colour scale represents the normalized intensity of Na at various locations of the meat cross section. Fig. 3 (a) illustrates the cross-section image of the control sample which clearly shows the emission intensities of salt (Na) equally distributed throughout the sample at the lowest level of the intensity colour scale. Fig. 3 (b) illustrates the cross-sectional image of the 2h brined sample, where higher intensities of Na were evident along the edges, which decreases towards the centre illustrating salt diffusion from all sides. Lower Na intensities were observed at the right edge of the sample as compared to other sides which could be related to non-uniformity of the sample, subsequently affecting the diffusion process. As expected, salt diffusion increased with brining time of 24 h which is evident from Fig. 3 (c). The variability in Na intensities along the edges could be caused by the same reason as for 2h brined sample. Ideally, the distribution of salt should be uniform from all sides; however factors such as sample geometry, sample size, brine concentration and brine temperature plays a significant role in defining the diffusion process (Chabbouh et al., 2012). Also factors affecting the LIBS spectra such as LTSD, sample uniformity and matrix cannot be neglected (Markiewicz-Keszycka et al., 2017; Radziemski and Cremers, 2006). Overall, it can be concluded that LIBS can be a rapid and chemical-free technique to image salt diffusion as well different minerals in various meats.

For future studies, factors affecting the LIBS spectra as well as salt diffusion for various meats should be studied. Reference values for salt or Na should be obtained in order to correlate the actual concentration values with spectral intensities.

4. Conclusions

This study illustrates the ability of LIBS as a novel spectral technique for imaging NaCl diffusion in beef by utilizing the Na emission peak at 589.05 nm in LIBS spectra. Spectral images of the cross-section of brined beef samples (2h and 24 h) along with a control sample were generated with respect to the Na emission peak at 589.05 nm. As expected, the control sample showed the lowest salt (Na) distribution whereas 2h brined sample showed Na distribution along the sample edges which decreased towards the centre illustrating salt diffusion. Salt diffusion increased for the 24 h brined sample, indicated by an increased gradient of Na intensity towards the centre. Variability in the Na intensities along the edges was observed which could be related to non-uniformity of the sample, subsequently

affecting the diffusion process. Overall, results illustrate the ability of LIBS combined with an automated chamber to map salt diffusion in meat via Na LIBS imaging. The findings presented here show the potential for LIBS in optimizing brining time, brine concentration as well as brine temperature.

5. Acknowledgements

The authors would like to acknowledge funding from the Food Institutional Research Measure administered by the Department of Agriculture, Food and the Marine, Ireland (Grant agreement: 13/F/508).

6. References

- Abdel-Salam, Z., Abdel-Salam, S., Harith, M., (2017). Application of Laser Spectrochemical Analytical Techniques to Follow Up Spoilage of White Meat in Chicken. *Food Analytical Methods* 10(7), 1-8.
- Andersen, M.-B.S., Frydenvang, J., Henckel, P., Rinnan, Å., (2016). The potential of laser-induced breakdown spectroscopy for industrial at-line monitoring of calcium content in comminuted poultry meat. *Food Control* 64, 226-233.
- Bilge, G., Sezer, B., Eseller, K.E., Berberoglu, H., Koksel, H., Boyaci, I.H., (2016a). Ash analysis of flour sample by using laser-induced breakdown spectroscopy. *Spectrochimica Acta Part B: Atomic Spectroscopy* 124, 74-78.
- Bilge, G., Velioglu, H.M., Sezer, B., Eseller, K.E., Boyaci, I.H., (2016b). Identification of meat species by using laser-induced breakdown spectroscopy. *Meat Sci* 119, 118-122.
- Cama-Moncunill, X., Markiewicz-Keszycka, M., Dixit, Y., Cama-Moncunill, R., Casado-Gavaldà, M.P., Cullen, P.J., Sullivan, C., (2017). Feasibility of Laser-Induced Breakdown Spectroscopy (LIBS) as an at-line validation tool for calcium determination in infant formula. *Food Control* 78, 304-310.
- Casado-Gavaldà, M.P., Dixit, Y., Geulen, D., Cama-Moncunill, R., Cama-Moncunill, X., Markiewicz-Keszycka, M., Cullen, P.J., Sullivan, C., (2017). Quantification of copper content with Laser Induced Breakdown Spectroscopy as a potential indicator of offal adulteration in beef. *Talanta* 169, 123-129.
- Castro, J.P., Pereira-Filho, E.R., (2016). Twelve different types of data normalization for the proposition of classification, univariate and multivariate regression models for the direct analyses of alloys by laser-induced breakdown spectroscopy (LIBS). *Journal of Analytical Atomic Spectrometry* 31(10), 2005-2014.

182 Chabbouh, M., Ahmed, S.B.H., Farhat, A., Sahli, A., Bellagha, S., (2012). Studies on the salting step of
 183 Tunisian Kaddid meat: Experimental kinetics, modeling and quality. Food and Bioprocess Technology
 184 5(5), 1882-1895.

185 Crank, J., (1979). *The mathematics of diffusion*. Oxford university press.

186 Dixit, Y., Cama, R., Sullivan, C., Alvarez Jubete, L., Ktenioudaki, A., (2014). Near infrared
 187 hyperspectral image analysis using R. Part 5: Animated visualisation of hyperspectral data using R and
 188 ImageJ. NIR news 25(7), 15-17.

189 Dixit, Y., Casado-Gavaldà, M., Cama-Moncunill, R., Cama-Moncunill, X., Markiewicz-Keszycka,
 190 M., Cullen, P., Sullivan, C., (2017a). Laser Induced Breakdown Spectroscopy for quantification of
 191 sodium and potassium in minced beef: a potential technique for detecting beef kidney adulteration.
 192 Analytical Methods 9(22), 3314-3322.

193 Dixit, Y., Casado-Gavaldà, M.P., Cama-Moncunill, R., Cama-Moncunill, X., Cullen, P., Sullivan, C.,
 194 (2017b). Quantification of rubidium as a trace element in beef using laser induced breakdown
 195 spectroscopy. Meat Sci 130, 47-49.

196 Er, S.V., Eksi-Kocak, H., Yetim, H., Boyaci, I.H., (2016). Novel Spectroscopic Method for
 197 Determination and Quantification of Saffron Adulteration. Food Analytical Methods 10(5), 1-9.

198 Gowen, A., (2014). Near infrared hyperspectral image analysis using R. Part 2: working with
 199 hypercubes. NIR news 25(3), 17-19.

200 Graiver, N., Pinotti, A., Califano, A., Zaritzky, N., (2006). Diffusion of sodium chloride in pork tissue.
 201 Journal of Food Engineering 77(4), 910-918.

202 Graiver, N., Pinotti, A., Califano, A., Zaritzky, N., (2009). Mathematical modeling of the uptake of
 203 curing salts in pork meat. Journal of Food Engineering 95(4), 533-540.

204 Hansen, C.L., van der Berg, F., Ringgaard, S., Stødkilde-Jørgensen, H., Karlsson, A.H., (2008).
 205 Diffusion of NaCl in meat studied by ^1H and ^{23}Na magnetic resonance imaging. Meat Sci 80(3), 851-
 206 856.

207 Lawrie, R.A.a.L., D.A., (2006). *Lawrie's meat science*. Woodhead Publishing Limited.

208 Markiewicz-Keszycka, M., Cama-Moncunill, X., Casado-Gavaldà, M.P., Dixit, Y., Cama-Moncunill,
 209 R., Cullen, P.J., Sullivan, C., (2017). Laser-induced breakdown spectroscopy (LIBS) for food analysis:
 210 A review. *Trends in Food Science & Technology* 65, 80-93.

211 Moncayo, S., Rosales, J.D., Izquierdo-Hornillos, R., Anzano, J., Cáceres, J.O., (2016). Classification of
 212 red wine based on its protected designation of origin (PDO) using Laser-induced Breakdown
 213 Spectroscopy (LIBS). *Talanta* 158, 185-191.

214 R Core Team, (2014). R: A language and environment for statistical computing. R Foundation for
 215 Statistical Computing, Vienna, Austria, 2012. ISBN 3-900051-07-0.

216 Radziemski, L.J., Cremers, D.A., (2006). *Handbook of Laser Induced Breakdown Spectroscopy*. John
 217 Wiley & Sons, West Sussex, England.

218 Rinnan, Å., van den Berg, F., Engelsen, S.B., (2009). Review of the most common pre-processing
 219 techniques for near-infrared spectra. *TrAC Trends in Analytical Chemistry* 28(10), 1201-1222.

220 Sharedeh, D., Mirade, P.-S., Venien, A., Daudin, J.-D., (2015). Analysis of salt penetration
 221 enhancement in meat tissue by mechanical treatment using a tumbling simulator. *Journal of Food*
 222 *Engineering* 166, 377-383.

223 Singh, J., Kumar, R., Awasthi, S., Singh, V., Rai, A., (2017). Laser Induced breakdown spectroscopy: A
 224 rapid tool for the identification and quantification of minerals in cucurbit seeds. *Food Chem* 221, 1778-
 225 1783.

226 Skou, J.C., (1988). [1] Overview: The Na, K-pump. *Methods in enzymology* 156, 1-25.

227 Wang, J., Zheng, P., Liu, H., Fang, L., (2016). Classification of Chinese tea leaves using laser-induced
 228 breakdown spectroscopy combined with the discriminant analysis method. *Analytical Methods* 8(15),
 229 3204-3209.

230 Weritz, F., Schaurich, D., Wilsch, G., (2007). Detector comparison for sulfur and chlorine detection
 231 with laser induced breakdown spectroscopy in the near-infrared-region. *Spectrochimica Acta Part B:*
 232 *Atomic Spectroscopy* 62(12), 1504-1511.

233